



# California Medical Device Recall Information



## Recall Name

### Bunnell Recalls Life Pulse High-Frequency Ventilator Patient Circuits Due to Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
12/12/12	Life Pulse High-Frequency Ventilator Patient Circuits	<b>Bunnell, Inc.</b> Salt Lake City, UT	<i>Found to have heater wire insulation that can melt, causing sparking and smoke</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Life Pulse High-Frequency Ventilator Patient Circuits  <a href="#">List of Suspect Lots Recalled</a>  ( <b>Note:</b> Do not reference the Lot # on the outside of the circuit box. The recalled patient circuit lot # can be found on a label attached to the humidifier cartridge of each individual circuit.)	<b>CA</b> , nationwide and Canada	Distributed since March 19, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm330842.htm>